



NATIONAL TYPE EVALUATION PROGRAM

Certificate of Conformance

for Weighing and Measuring Devices

For:

Non-Computing Scale
Bench Scale, Load Cell Electronic
Model: EB Series
 n_{max} : 3 000
 e_{min} : (see below)
Capacity: (see below)
Platform: (see below)
Accuracy Class: III

***Submitted By: Contact Info. Updated: October 2010**

Cardinal Scale Manufacturing Co.
203 East Daugherty
Webb City, MO 64870
Tel: 417-673-4631
Fax: 417-673-5001
Contact: Stephen Langford
Email: slangford@cardet.com
Web site: www.cardinalscales.com

Standard Features and Options

XXX = Indicator Type:

- 204 Cardinal Model 204 Weight Indicator (NTEP CC 02-016)
- 205 Cardinal Model 205 Weight Indicator (NTEP CC 01-011)
- 210 Cardinal Model 210 Weight Indicator (NTEP CC 01-011)
- LPAN No Indicating Element
- Construction: Mild Steel


Load Cell Used:


- Cardinal TSP Series strain gauge single point load cells (NTEP CC 97-136) and Cardinal TSSP-200KG single point strain gauge load cell (Model EB-300-XXX only) or approved metrological equivalent.

Model No.	Capacity	Scale Interval	Platform Size	Load Cell
EB-15-XXX	15 lb / 6 kg	0.005 lb / 0.002 kg	10 x 12 in / 250 x 300 mm	Cardinal TSP-10 kg
EB-30-XXX	30 lb / 15 kg	0.01 lb / 0.005 kg	10 x 12 in / 250 x 300 mm	Cardinal TSP-20 kg
EB-60-XXX	60 lb / 30 kg	0.02 lb / 0.01 kg	14 x 16 in / 350 x 400 mm	Cardinal TSP-50 kg
EB-150-XXX	150 lb / 60 kg	0.05 lb / 0.02 kg	14 x 16 in / 350 x 400 mm	Cardinal TSP-100 kg
EB-300-XXX	300 lb / 150 kg	0.1 lb / 0.05 kg	20 x 24 in / 500 x 600 mm	Cardinal TSP-200 kg

Temperature Range: -10 °C to 40 °C (14 °F to 104 °F)

This device was evaluated under the National Type Evaluation Program and was found to comply with the applicable technical requirements of "NIST Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices." Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages. *Editorial changes, not affecting the type or metrological content, corrected this certificate.


Tim Tyson
Chairman, NCWM, Inc.


Randy Jennings
Chairman, National Type Evaluation Program Committee
Issued: October 19, 2010

1135 M Street, Suite 110 / Lincoln, Nebraska 68508

The National Conference on Weights and Measures (NCWM) does not approve, recommend or endorse any proprietary product or material, either as a single item or as a class or group. Results shall not be used in advertising or sales promotion to indicate explicit or implicit endorsement of the product or material by the NCWM.



Cardinal Scale Manufacturing Co.

Non-Computing Scale / EB Series

Application: General purpose load-receiving element when connected to an approved and compatible indicating element.

Identification: The required information is on a label located on top of the weighbridge. Access to this identification label may be gained by lifting upward on the stainless steel top cover and removing it from its support.

Sealing: There are no metrological adjustments in the load-receiving element. All calibration adjustments take place within the indicating element.

Test Conditions: This Certificate supersedes Certificate of Conformance number 03-031 and was issued to add the Models EB-15-XXX and EB-30-XXX bench scales. The Model EB-15-LPAN, 15 x 0.005 lb (6 x 0.002 kg) was submitted for this evaluation. The device was interfaced with a Cardinal Model 210 (Certificate of Conformance Number 01-011A1) indicating element. The emphasis of the evaluation was on the device design, marking, operation and compliance with influence factor requirements. Several increasing/decreasing load and shift tests were performed. The device was tested over a temperature range of -10 °C to 40 °C (14 °F to 104 °F). A load of half capacity was applied to the base over 100 000 times, the scale was tested periodically during this time. Previous test conditions are listed below for reference.

Certificate of Conformance Number 03-031: The Model EB-60-LPAN, 60 x 0.02 lb (30 x 0.01 kg) bench scale and the Model EB-300-LPAN, 300 x 0.1 lb (150 x 0.05 kg) bench scale were submitted for this evaluation. The devices were interfaced with a Cardinal Model 210 (NTEP Certificate of Conformance Number 01-011A1) indicating element. The emphasis of the evaluation was on device design, markings, operation and compliance with influence factor requirements. Several increasing/decreasing load and shift tests were performed. The device was tested over a temperature range of -10 °C to 40 °C (14 °F to 104 °F). A load of half capacity was applied to the base over 100 000 times, the scale was tested periodically during this time.

Evaluated By: A. McCoy (OH) 03-031 and 03-031A1

Type Evaluation Criteria Used: NIST, Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices, 2003. NCWM, Publication 14: Weighing Devices, 2003.

Conclusion: The results of the evaluation and information provided by the manufacturer indicate the device complies with applicable requirements.

Information Reviewed By: S. Patoray (NCWM), L. Bernetich (NCWM) 03-031, 03-031A1