



**CE DECLARATION OF CONFORMITY**

<b>Instrument Model:</b>	ID-3636##-855###, ID-3636##-855###-LPTN, ID-4836##-855###, ID-4836##-855###-LPTN, ID-4848##-855###, ID-4848##-855###-LPTN, ID-7248##-855###, ID-7248##-855###-LPTN <b>In Floor Dialysis Scale</b>
<b>Manufacturer:</b>	Cardinal Scale Manufacturing Company Detecto Scale 203 East Daugherty Street Webb City, Missouri 64870 USA Telephone No. + 417 673 4631 Fax No. + 417 673 5001
<b>Type Approval Certificate(s):</b>	855 DK0199.583 LFB500-HMT R60/1991-DK-00.02
<b>Test Certificate(s):</b>	855 DANAK-1916732 LFB500-HMT DANAK-193262
<b>Notified Body Issuing Approval for Module B: EU Type Examination (Annex II.2 of 2014/31/EU):</b>	0199-Delta

Conformity to the following Directives is demonstrated by reference to the following harmonized standards or technical documentation

<u>Applicable Directives</u>	<u>Harmonized Standards or Other Technical Documentation</u>
<b>Directive on Non-Automatic Weighing Instruments (2014/31) including all Essential Requirements of Annex I:</b>	EN45501:2015/Module B
<b>Directive on Electromagnetic Compatibility/Low Voltage (2014/35):</b>	EN60950-1:2005 (Second Edition) + Am 1:2009 + Am 2:2013
<b>Directive for Medical Devices, 93/42/EU Class I Medical Device</b>	EN ISO13485:2012 Certificate 1004213 issued by Orion Registrar, Inc.

The Technical Construction File required by this Directive is maintained at the corporate headquarters of Cardinal Scale Manufacturing Company.

This declaration is issued under the sole responsibility of the manufacturer.

Signed for and on behalf of Cardinal Scale Manufacturing 203 East Daugherty, Webb City MO, 64870:

Mark Levels  
Manager, Quality Assurance

6/29/2017

Rev A

EU-ID-KGEU