

# TEST CERTIFICATE

## No. DK 0199.R60.8 Addition 1

**Issued by** DELTA Danish Electronics, Light & Acoustics,  
EC - Notified Body No. 0199.

**In accordance with** EN 45501 (1992), paragraph 8.1 and 3.5.4  
with fraction  $p_{LC} = 0.7$ .  
OIML R60 (Edition 1991).

**Issued to** **Cardinal Scale Manufacturing Co.**  
203 East Daugherty St.  
Webb City, Missouri 64870  
U.S.A.

**In respect of** Single point, bending beam, strain gauge load cell.

**Type** TSP.

**Manufacturer** **Cardinal Scale Manufacturing Co.**  
203 East Daugherty St.  
Webb City, Missouri 64870  
U.S.A.

**Addition 1** 40 kg and 80 kg capacities are included

### Characteristics

Classification		C3	C4
Maximum number of LC verification intervals	$n_{LC}$	3000	4000
Maximum capacity in kg	$E_{max}$	10, 15, 20, 30, 40, 50, 80, 100	
Minimum LC verification interval	$v_{min}$	$E_{max} / 10000$	
Minimum dead load output return	MDLOR	0.017 % of $E_{max}$	
Temperature limits		- 10 °C / + 40 °C	

### DELTA

Danish Electronics,  
Light & Acoustics

Venlighedsvej 4  
2970 Hørsholm  
Denmark

Tel. (+45) 72 19 40 00  
Fax (+45) 72 19 40 01  
www.delta.dk

**Issued on** 2003-07-16

**Signatory :**   
**P. Bengtson**

